



ULTRA FOR MEDICAL PRODUCTS CO.

INSTRUCTION FOR USE FOR

ULTRAMED LATEX FOLEY BALLON CATHETER

Material Used:

- Latex Foley catheters: Made of latex and the surface of catheter is silicon coated.

Indications:

- To drain the bladder in case of acute urinary retention.
- For post operative drainage and irrigation of bladder.

Contraindications:

- Products that contain latex may cause allergic reaction.
- Long term complication; Stricture of the urethra.
- Standard Medical text books and scientific literature should be consulted for specific adverse reaction information.

Instruction for Use:

(In all instances aseptic technique must be used)

Inflation Procedure

- Insert the luer tip of the syringe with a firm push/twist motion into the catheter valve housing to assure activation of the valve.
- Inflate with the prescribed volume i.e. 1.5 ml/cc balloon: maximum 1.5 ml/cc sterile water; 3 ml/cc balloon: maximum 5 ml/cc sterile water; 5/10 ml/cc balloon: maximum 10 ml/cc sterile water; 30 ml/cc balloon: maximum 30 ml/cc sterile water; 50 ml/cc balloon: maximum 50 ml/cc sterile water.
- While holding syringe plunger check for any evidence of leaking or deflation.
- Release syringe plunger and allow balloon to deflate and observe for any difficulty with the deflation. Use only gentle aspiration, if necessary, to deflate balloon.
- Carefully insert the catheter shaft into the urethral meatus until its tip and balloon are suitably located in the bladder, normally indicated by urine flow. Then slowly inflate the balloon with the prescribed volume of sterile water and gently pull the catheter shaft backwards ensuring that the balloon is correctly "seated" at the neck of the bladder.

Deflation Procedure

- Insert the syringe luer tip into the valve housing with a firm push/twist motion.
- Release syringe plunger and allow balloon to deflate.
- Use only gentle aspiration, if necessary, to deflate balloon.
- The balloon should spontaneously deflate.
- If deflation does not occur, proceed to the alternative methods listed below.
 - STEP 1: Aspirate by slowly pulling back on syringe plunger handle to force deflation. If this does not work, proceed to Step 2.
 - STEP 2: Remove syringe and fill syringe with 2-3 ml/cc of sterile water; reinsert syringe tip into valve; inject water into balloon and aspirate. If this does not work, proceed to Step 3.
 - STEP 3: Use scissors to sever the side arm below the valve or cut the catheter shaft at some safe distance from the meatus and catheter should drain spontaneously. If this does not work, proceed to Step 4.
 - STEP 4: Inject 100-200 ml/cc of suitable irrigating fluid into bladder. Then slowly inject 10-15 ml/cc of mild mineral oil/water mixture into the balloon by syringe/needle and wait for deflation followed by a volume of a suitable irrigation fluid to avoid any chance of chemical cystitis. If this does not work, proceed to Step 5.
 - STEP 5: Scope around the catheter and use a suitable device to puncture the balloon.



ULTRA FOR MEDICAL PRODUCTS CO.

Cautions:

- This product contains natural rubber latex which may cause allergic reaction.
- Do not clamp or kink catheter shaft. Clamping may damage catheter, preventing deflation. Use of sterile catheter spigot is recommended to stop urine flow.
- Do not use needle to inflate balloon.
- Do not aspirate urine through the drainage funnel wall.
- Use water soluble lubricants only. Do not use ointments or lubricants having a petrolatum base. They will damage catheter's material and may burst balloon.
- Fill catheter balloon with sterile water only.
- Do not expose silicone catheter to radiation as it will change the properties of the silicone rubber to those caused by heat ageing.
- Do not use if package is open or damaged.
- Discard after single use.

Conditions of Handling, Preservation and Storage:

- Not more than 5 cartoons on each other.
- Nice Ventilated place.
- Out of Sun light.
- UV lights and high dosage of radiation should be avoided during storage.
- Ozone is deleterious to latex catheter. The storage environment should not contain any equipment that generates ozone, such as mercury vapour lamps, high voltage electrical appliances or others that give rise to electric sparks or silent electrical discharges.
- Latex catheter should be protected from metals such as copper, manganese and iron.
- Room Temperature.



Obelis^{SA}

53 Bd. Général Wahis
B-1030 Brussels, Belgium
Phone: 32.2.732.59.54
Fax: 32.2.732.60.03
E-mail: mail@obelis.net
www.obelis.net



Sterilized by Ethylene Oxide



For Single Use Only



Cautions



Lot No.



Date of Manufacturing



Use by / Expiry date



Medical Device



Non-Pyrogenic



Catalogue No.



Manufacturer Address



Don't use if package is not intact



Latex



Consult Instruction for Use



Non-phthalate



ISSUE/REV DATE: 01.09.2022

Signature
Astrea

Assiut Factory: Part No. (304:310) & part no. (312) – Industrial Area, Arab El Awamer – Abnoub - Assiut – Egypt. Tel: 002-088-4964333 (500), 002-088-4964666 (600), 002-088-4964777 (700), 002-088-4964888 (800), 002-088-4964999 (900)

& Fax: 002-088-4964222 & **Mob:** 002-01001558853, 002-01068832355

Cairo Head Office: 64, Nakhla El Motaiy Triumph Square Heliopolis Tel: 022/4171621-4143794 & Fax: 022/4171613 & Mob: 01223988200

Assiut Office: 23, July Str. Tel: 088/2364111 – 2364222 & Fax: 088/2334964 & Mob. : 01223988202

Alexandria Office: 212 Abd El salaam Airef Str, Luran Tel: 03/5856202– 5856458 & Fax: 03/5828988 & Mob: 01223948666

E-mail: ultra@elaggargroup.com, shady@elaggargroup.com

Website: www.ultramedumic.com

Instruction For Use

ISSUE/REV: - 2/0

Page 2 of 2

FOLEY CATHETER